## **MPOR** instrument series



Pocket Instruments with USB-Interface for Convenient and Fast Coating Thickness Measurement on Virtually all Metals







## MPOR instrument series - Overview

Overview of the various models of the MPOR instrument series

	Application		Probe	Probe	USB	
Instrument models	NC/Fe	NF/Fe	NC/NF	integrated in the instru- ment case	with cable perma- nently connected at the instrument	Mini USB port
PERMASCOPE® MPOR 605-117, see page 3	•	•		•		•
PERMASCOPE® MPOR-FP 605-118, see page 3	•	•			•	•
ISOSCOPE® MPOR 605-116, see page 7			•	•		•
DUALSCOPE® MPOR 605-097, see page 11	•	•	•	•		•
DUALSCOPE® MPOR-FP 605-114, see page 11	•	•	•		•	•
DUALSCOPE® MPORH-FP 605-115, see page 15	•	•	•		•	•

NC/Fe: Non conductive coating material on ferrous metals NF/Fe: Non magnetic coating material on ferrous metals NC/NF: Non conductive coating material on non-ferrous metals

2 MPOR instrument series

# PERMASCOPE® MPOR PERMASCOPE® MPOR-FP

Pocket Instruments with PC-Interface for Convenient and Fast Coating Thickness Measurement on Steel and Iron





# PERMASCOPE® MPOR Models

Description	
	The PERMASCOPE measuring instruments measure coating thicknesses easily, quickly, non-destructively and with the precision that is typical for all Fischer instruments.
Instrument properties	<ul> <li>Ideal for onsite applications due to the compact size, the light weight and the robust and durable instrument design</li> </ul>
	<ul> <li>Intuitive operation of the menu navigation and graphic display. The display turns auto- matically, like a smart phone</li> </ul>
	<ul> <li>Second display for reading the measurement results directly on the top side of the instrument, e.g., for measuring overhead</li> </ul>
	<ul><li>Different languages are selectable</li><li>Manufacturer's certificate, included in the scope of supply</li></ul>
Generating measurements	<ul> <li>The specimen's shape and permeability have a comparatively low influence on the measurement results</li> </ul>
	<ul> <li>Two special measuring modes in accordance with the measurement regulations IMO PSPC (90/10-Rule) and SSPC-PA2</li> </ul>
Applications	Steel or iron substrates (Fe)
Examples	<ul> <li>Zinc, chromium, copper, paint, varnish and plastic coatings on steel, iron or cast iron (Fe)</li> </ul>
	<ul> <li>Measurements both on smooth and rough surfaces</li> </ul>
	The instruments are particularly suited for highly precise measurements of thin coating.
Models	
	<ul> <li>PERMASCOPE MPOR: Probe integrated in the measuring instrument for single-handed operation</li> </ul>
	<ul> <li>PERMASCOPE MPOR-FP: Probe with cable (80 cm; 31.5 ") permanently connected to the instrument, for measurements on various specimen shapes</li> </ul>
Evaluation	
Statistics	Display of mean value, standard deviation, MIN, MAX and number of measurements per block
PC software included in the scope of supply	PC software FISCHER DataCenter with the following functionality: Transferring and archiving measurement data, comprehensive statistical and graphical evaluations, easy creation and printing of inspection reports
Measuring Modes	
Standard (Std)	Standard measuring mode for simple, universal coating thickness measurements, all measurement functions are available.
IMO PSPC 90/10 (90.10)	90/10 rule stored in the instrument for coating thickness measurements according to the requirements of the "Performance Standard for Protective Coatings" of the International Maritime Organization (IMO PSPC).
SSPC-PA2 (SSPC)	Coating thickness measurement according to the test specification SSPC-PA2 of the Society for Protective Coatings (SSPC).

Block size Adjustable between 2 and 20 single readings per block

Tolerance limits Adjustable, depending on the selected measuring mode

Offset value In the standard mode, the freely adjustable offset value is deducted automatically from the

measured value. Thus, one obtains the thickness of the top coating if for instance the

interim coating is known.

Units of measurement Selectable µm or mils

Continuous display mode Measurement in "continuous display mode" for continuous sampling of the surfaces, e.g.,

in the manufacture of tanks and containers.

Normalization Adaptation to the substrate material and the shape of the specimen.

Factory calibration

Each individual instrument is factory calibrated at several reference points with the great-

est care to ensure the highest possible degree of trueness.

Corrective calibration (Adjustment)

Adaptation to the substrate material and the shape of the specimen and to a thickness

value using a calibration foil.

Simple Calibration

Adaption to the coating and substrate material in one step using a coated reference part with a coating thickness higher than 200 µm (7.87 inches). Nevertheless, this kind of calibration supplies only a lower accuracy as specified in the sections Trueness and Repeata-

bility Precision.

#### **General Features**

Calibration

Measuring method Magnetic induction method (ISO 2178, ASTM D7091, Measurement of non-magnetic

coatings on magnetic substrates)

Probe tip radius: 2 mm (78 mils); Probe tip material: Hard metal

Data memory Max. 10,000 individual readings; the contents of the memory is retained even without bat-

teries

Measuring frequency More than 70 measurements per minute

Measurement acquisition Automatic upon placement of the probe; indication of the measurement with a beep visu-

ally with a green lit LED

Display limit value violation Acoustically through 2 short beeps and visually with a red lit LED

Display

• Graphic display with an automatically turning display in order to read the measurement

results in many different instrument positions

• LCD display on the top side of the instrument, e.g., for reading the measurement value

for measuring overhead

Languages Many different display languages are selectable: German, English and several other

European and Asian languages

USB port 2.0 compatible, mini type B socket, for connecting a PC

Data transfer Single readings, mean values, group separator

Admissible ambient temperature

range during operation

0 ... +40 °C (+32 ... +104 °F)

PERMASCOPE® MPOR Models 5

### PERMASCOPE® MPOR Models

Weight (incl. batteries) MPOR: 137 g (4.8 oz)


MPOR-FP: 184 g (6.5 oz)

Power supply 2 Batteries, LR6, AA, 1.5 V

#### **Dimensions**

Instrument Width: 64 mm (2.5 "); depth: 28 mm (1.1 "); height: 85 mm (3.35 ")

Probe of instrument MPOR-FP



#### **Measurement Range**

0 ... 2500 µm (97.5 mils)

#### **Trueness**

based on factory calibration stand- $0 \dots 3.9 \text{ mils} \le 0.06 \text{ mils}$  $0 \dots 100 \ \mu m \le 1.5 \ \mu m$ ards of the Helmut Fischer GmbH  $3.9...39 \text{ mils:} \leq 1.5\% \text{ of reading}$  $100 \dots 1000 \ \mu m \le 1.5 \%$  of reading  $39 \dots 97.5 \text{ mils} : \leq 3 \% \text{ of reading}$ 1000 ... 2500  $\mu m$ :  $\leq$  3 % of reading

**Repeatability Precision** 

 $0 \dots 100 \ \mu m \le 0.3 \ \mu m$ ...  $3.9 \text{ mils}: \le 0.0117 \text{ mils}$ based on factory calibration standards of the Helmut Fischer GmbH,  $100 \dots 2500 \ \mu m \le 0.3 \% \ of \ reading$  $3.9...97.5 \text{ mils: } \le 0.3\% \text{ of reading}$ 5 single measurement readings on

**Ordering Data** 

each standard

PERMASCOPE MPOR, probe integrated in the measuring instrument 605-117 605-118 PERMASCOPE MPOR-FP, probe with cable (80 cm; 31.5 ") permanently connected to the instrument

#### **Scope of Supply**

Instrument case; protective instrument cover; lanyard; 2 batteries; metal plate NF/FE for testing purposes; calibration foil (foil thickness about 75 µm (2.95 inches)); operator's manual; manufacturer's certificate; USB cable; support CD with USB drivers, software program FISCHER DataCenter for convenient evaluating, documenting and archiving of the measurement data, software program PC-Datex for exporting the measurement data to an Excel spreadsheet

PERMASCOPE® is a registered trademark of Helmut Fischer GmbH Institut für Elektronik und Messtechnik in Germany and in other countries.







# ISOSCOPE® MPOR

Pocket Instrument with PC-Interface for Convenient and Fast Coating Thickness Measurement on Virtually all Non-Ferrous Metals





# ISOSCOPE® MPOR

Description				
	The ISOSCOPE measuring instrument measures coating thicknesses easily, quickly, non-destructively and with the precision that is typical for all Fischer instruments.			
Instrument properties	<ul> <li>Ideal for onsite applications due to the compact size, the light weight and the robust and durable instrument design</li> </ul>			
	<ul> <li>Intuitive operation of the menu navigation and graphic display. The display turns auto- matically, like a smart phone</li> </ul>			
	<ul> <li>Second display for reading the measurement results directly on the top side of the instrument, e.g., for measuring overhead</li> </ul>			
	Different languages are selectable			
	<ul> <li>Manufacturer's certificate, included in the scope of supply</li> </ul>			
Generating measurements	<ul> <li>The specimen's shape and permeability have a comparatively low influence on the measurement results</li> </ul>			
	<ul> <li>Patented conductivity compensation for measurements on non-magnetic substrate materials</li> </ul>			
	<ul> <li>Two special measuring modes in accordance with the measurement regulations IMO PSPC (90/10-Rule) and SSPC-PA2</li> </ul>			
Applications	Nonferrous metal substrates (NF)			
Examples	Paint, varnish or plastic coatings on aluminium, copper or brass			
	Anodized coatings on aluminium			
	The instrument is particularly suited for highly precise measurements of thin coatings.			
Evaluation				
Statistics	Display of mean value, standard deviation, MIN, MAX and number of measurements per block			
PC software	PC software FISCHER DataCenter with the following functionality: Transferring and			
included in the scope of supply	archiving measurement data, comprehensive statistical and graphical evaluations, easy creation and printing of inspection reports			
Measuring Modes				
Standard (Std)	Standard measuring mode for simple, universal coating thickness measurements, all measurement functions are available.			
IMO PSPC 90/10 (90.10)	90/10 rule stored in the instrument for coating thickness measurements according to the requirements of the "Performance Standard for Protective Coatings" of the International Maritime Organization (IMO PSPC).			
SSPC-PA2 (SSPC)	Coating thickness measurement according to the test specification SSPC-PA2 of the Society			

for Protective Coatings (SSPC).

Block size Adjustable between 2 and 20 single readings per block

Tolerance limits Adjustable, depending on the selected measuring mode

Offset value In the standard mode, the freely adjustable offset value is deducted automatically from the

measured value. Thus, one obtains the thickness of the top coating if for instance the

interim coating is known.

Units of measurement Selectable µm or mils

Continuous display mode Measurement in "continuous display mode" for continuous sampling of the surfaces, e.g.,

in the manufacture of tanks and containers.

Normalization Adaptation to the substrate material and the shape of the specimen.

Calibration Factory calibration

Each individual instrument is factory calibrated at several reference points with the great-

Lach individual instrument is factory calibrated at several reference points with the grec

est care to ensure the highest possible degree of trueness. Corrective calibration (Adjustment)

Adaptation to the substrate material and the shape of the specimen and to a thickness

value using a calibration foil.

Simple Calibration

Adaption to the coating and substrate material in one step using a coated reference part with a coating thickness higher than 200 µm (7.87 inches). Nevertheless, this kind of calibration supplies only a lower accuracy as specified in the sections Trueness and Repeata-

bility Precision.

#### **General Features**

Measuring method Eddy current method (ISO 2360, ASTM D7091, Measurement of non-conductive coatings

on non-magnetic substrate metals)

Probe Probe tip radius: 1.2 mm (46.8 mils); Probe tip material: Ruby

Data memory Max. 10,000 individual readings; the contents of the memory is retained even without

batteries

Measuring frequency More than 70 measurements per minute

Measurement acquisition Automatic upon placement of the probe; indication of the measurement with a beep

visually with a green lit LED

Display limit value violation Acoustically through 2 short beeps and visually with a red lit LED

Display

• Graphic display with an automatically turning display in order to read the measurement

results in many different instrument positions

• LCD display on the top side of the instrument, e.g., for reading the measurement value

for measuring overhead

Languages Many different display languages are selectable: German, English and several other

European and Asian languages

USB port 2.0 compatible, mini type B socket, for connecting a PC

Data transfer Single readings, mean values, group separator

Admissible ambient temperature

range during operation

0 ... +40 °C (+32 ... +104 °F)

ISOSCOPE® MPOR

## ISOSCOPE® MPOR

Weight (incl. batteries) 137 g (4.8 oz)

Power supply 2 Batteries, LR6, AA, 1.5 V

Width: 64 mm (2.5 "); depth: 28 mm (1.1 "); height: 85 mm (3.35 ") Dimensions (W  $\times$  D  $\times$  H)

#### **Measurement Range**

0 ... 1200 µm (46.8 mils)

#### **Trueness**

based on factory calibration stand-0 ...  $70~\mu m$ :  $\leq 1.0~\mu m$ ... 2.7 mils:  $\leq 0.039 \text{ mils}$ ards of the Helmut Fischer GmbH 70 ... 250  $\mu m$ :  $\leq 1.5 \%$  of reading 2.7 ... 9.75 mils:  $\leq 1.5 \%$  of reading  $250 \dots 1000 \, \mu m \le 3 \% \text{ of reading}$  $9.75 \dots 39 \text{ mils}$ :  $\leq 3 \% \text{ of reading}$ 

#### **Repeatability Precision**

based on factory calibration stand- $50 \ \mu m$ :  $\leq 0.25 \ \mu m$  $0 \dots 2 \text{ mils} \le 0.0098 \text{ mils}$ 0 ...  $2 \dots 39 \text{ mils} \le 0.5 \% \text{ of reading}$ ards of the Helmut Fischer GmbH,  $50 \dots 1000 \; \mu m \colon \leq 0.5 \; \%$  of reading 5 single measurement readings on

#### **Ordering Data**

each standard

605-116 ISOSCOPE MPOR, probe integrated in the measuring instrument

#### Scope of Supply

Instrument case; protective instrument cover; lanyard; 2 batteries; metal plate ISO/NF for testing purposes; calibration foil (foil thickness about 75 µm (2.95 inches)); operator's manual; manufacturer's certificate; USB cable; support CD with USB drivers, software program FISCHER DataCenter for convenient evaluating, documenting and archiving of the measurement data, software program PC-Datex for exporting the measurement data to an Excel spreadsheet

ISOSCOPE® is a registered trademark of Helmut Fischer GmbH Institut für Elektronik und Messtechnik in Germany and in other countries.



# DUALSCOPE® MPOR DUALSCOPE® MPOR-FP

Pocket Instruments with PC-Interface for Convenient and Fast Coating Thickness Measurement on Virtually all Metals





# DUALSCOPE® MPOR Models

Description					
	The DUALSCOPE MPOR and MPOR-FP measuring instruments measure coating thicknesse easily, quickly, non-destructively and with the precision that is typical for all Fischer instruments.				
Instrument properties	<ul> <li>Ideal for onsite applications due to the com durable instrument design</li> </ul>	<ul> <li>Ideal for onsite applications due to the compact size, the light weight and the robust and durable instrument design</li> </ul>			
	<ul> <li>Intuitive operation of the menu navigation matically, like a smart phone</li> </ul>	and graphic display. The display turns auto-			
	ment, e.g., for measuring overhead  • Different languages are selectable				
	• Manufacturer's certificate, included in the	scope of supply			
Generating measurements	<ul> <li>The specimen's shape and permeability has measurement results</li> </ul>				
	materials				
	<ul> <li>Two special measuring modes in accordance</li> <li>IMO PSPC (90/10-Rule) and SSPC-PA2</li> </ul>	nce with the measurement regulations			
Applications	Steel or iron substrates (Fe)	Nonferrous metal substrates (NF)			
Examples	<ul> <li>Zinc, chromium, copper, paint, varnish and plastic coatings on steel, iron or cast</li> </ul>	<ul> <li>Paint, varnish or plastic coatings on aluminium, copper or brass</li> </ul>			
	iron (Fe)	Anodized coatings on aluminium			
	The instruments are applicable for measurem	nents both on smooth and rough surfaces			
Models					
	<ul> <li>DUALSCOPE MPOR: Probe integrated in the operation</li> </ul>	ne measuring instrument for single-handed			
	<ul> <li>DUALSCOPE MPOR-FP: Probe with cable (8 instrument, for measurements on various s</li> </ul>	80 cm; 31.5 ") permanently connected to the pecimen shapes			
Evaluation					
Statistics	Display of mean value, standard deviation, I	MIN, MAX and number of measurements per			
PC software	PC software FISCHER DataCenter with the fo	ollowing functionality: Transferring and			
included in the scope of supply	archiving measurement data, comprehensive creation and printing of inspection reports	archiving measurement data, comprehensive statistical and graphical evaluations, easy			
Measuring Modes					
Standard (Std)	Standard measuring mode for simple, univer measurement functions are available.	rsal coating thickness measurements, all			
IMO PSPC 90/10 (90.10)	90/10 rule stored in the instrument for coating thickness measurements according to the requirements of the "Performance Standard for Protective Coatings" of the International Maritime Organization (IMO PSPC).				
SSPC-PA2 (SSPC)	Coating thickness measurement according to the test specification SSPC-PA2 of the Society for Protective Coatings (SSPC).				

Block size Adjustable between 2 and 20 single readings per block

Tolerance limits Adjustable, depending on the selected measuring mode

Offset value In the standard mode, the freely adjustable offset value is deducted automatically from the

measured value. Thus, one obtains the thickness of the top coating if for instance the

interim coating is known.

Units of measurement Selectable µm or mils

Continuous display mode Measurement in "continuous display mode" for continuous sampling of the surfaces, e.g.,

in the manufacture of tanks and containers.

Normalization Adaptation to the substrate material and the shape of the specimen.

Factory calibration

Each individual instrument is factory calibrated at several reference points with the

greatest care to ensure the highest possible degree of trueness.

Corrective calibration (Adjustment)

Adaptation to the substrate material and the shape of the specimen and to a thickness

value using a calibration foil.

Simple Calibration

Adaption to the coating and substrate material in one step using a coated reference part with a coating thickness higher than 200 µm (7.87 inches). Nevertheless, this kind of calibration supplies only a lower accuracy as specified in the sections Trueness and Repeata-

bility Precision.

#### **General Features**

Calibration

Measuring method Magnetic induction method (ISO 2178, ASTM D7091, Measurement of non-magnetic

coatings on magnetic substrates);

Eddy current method (ISO 2360, ASTM D7091, Measurement of non-conductive coatings

on non-magnetic substrate metals);

Automatic selection of the measuring method corresponding to the substrate material

Probe tip radius: 2 mm (78 mils); Probe tip material: Hard metal

Data memory Max. 10,000 individual readings; the contents of the memory is retained even without

batteries

Measuring frequency More than 70 measurements per minute

Measurement acquisition Automatic upon placement of the probe; indication of the measurement with a beep

visually with a green lit LED

Display limit value violation Acoustically through 2 short beeps and visually with a red lit LED

Display

• Graphic display with an automatically turning display in order to read the measurement

results in many different instrument positions

• LCD display on the top side of the instrument, e.g., for reading the measurement value

for measuring overhead

Languages Many different display languages are selectable: German, English and several other

European and Asian languages

USB port 2.0 compatible, mini type B socket, for connecting a PC

Data transfer Single readings, mean values, group separator

DUALSCOPE® MPOR Models 13

## DUALSCOPE® MPOR Models

Admissible ambient temperature range during operation

0 ... +40 °C (+32 ... +104 °F)

Weight (incl. batteries)

MPOR: 137 g (4.8 oz); MPOR-FP: 184 g (6.5 oz)


Power supply

2 Batteries, LR6, AA, 1.5 V

#### **Dimensions**

Instrument Width: 64 mm (2.52 "); Depth: 28 mm (1.10 "); Height: 85 mm (3.35 ")

Probe of instrument MPOR-FP



Measurement Range	Steel or iron substrates (Fe)	Nonferrous metal substrates (NF)		
	0 2000 μm (78 mils)	0 2000 µm (78 mils)		
Trueness	Steel or iron substrates (Fe)	Nonferrous metal substrates (NF)		
based on factory calibration stand- ards of the Helmut Fischer GmbH	0 75 $\mu$ m: $\leq 1.5 \mu$ m 75 1000 $\mu$ m: $\leq 2$ % of reading 1000 2000 $\mu$ m: $\leq 3$ % of reading 0 2.9 mils: $\leq 0.06$ mils 2.9 39 mils: $\leq 2$ % of reading 39 78 mils: $\leq 3$ % of reading	0 50 $\mu$ m: $\leq$ 1 $\mu$ m 50 1000 $\mu$ m: $\leq$ 2 % of reading 1000 2000 $\mu$ m: $\leq$ 3 % of reading 0 2 mils: $\leq$ 0.039 mils 2 39 mils: $\leq$ 2 % of reading 39 78 mils: $\leq$ 3 % of reading		
Repeatability Precision	Steel or iron substrates (Fe)	Nonferrous metal substrates (NF)		
based on factory calibration stand- ards of the Helmut Fischer GmbH,	0 50 μm: ≤ 0.25 μm 50 2000 μm: ≤ 0.5 % of reading	0 100 μm: ≤ 0.5 μm 100 2000 μm: ≤ 0.5 % of reading		
5 single measurement readings on each standard	$0 \dots 2 \text{ mils:} \le 0.0098 \text{ mils}$ $2 \dots 78 \text{ mils:} \le 0.5 \% \text{ of reading}$	0 3.9 mils: ≤ 0.0195 mils 3.9 78 mils: ≤ 0.5 % of reading		
Ordering Data				
405.007	DIJAISCOPE AAROR washe intermented in the magnetic instrument			

605-097 DUALSCOPE MPOR, probe integrated in the measuring instrument

605-114 DUALSCOPE MPOR-FP, probe with cable (80 cm; 31.5 ") permanently connected to the

instrument

#### **Scope of Supply**

Instrument case; protective instrument cover; lanyard; 2 batteries; metal plates NF/FE and ISO/NF for testing purposes; calibration foil (foil thickness about 75 µm (2.95 inches)); operator's manual; manufacturer's certificate; USB cable; support CD with USB drivers, software program FISCHER DataCenter for convenient evaluating, documenting and archiving of the measurement data, software program PC-Datex for exporting the measurement data to an Excel spreadsheet

 $DUALSCOPE^{\circledR} is a \ registered \ trademark \ of \ Helmut \ Fischer \ GmbH \ Institut \ f\"ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ countries of \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ countries of \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ countries of \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ countries of \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ countries of \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ countries of \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ countries of \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ Germany \ and \ in \ other \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ Messtechnik \ in \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \ f\'ur \ Elektronik \ und \ GmbH \ Institut \$ 











# **DUALSCOPE® MPORH-FP**

Pocket Instrument with PC-Interface for Convenient and Fast Coating Thickness Measurement, especially on Thick Metal Coatings or Protective Coatings on Steel and Iron





# DUALSCOPE® MPORH-FP

Description						
	The DUALSCOPE MPORH-FP measuring instrument measures thick coatings easily, quick non-destructively and with the precision that is typical for all Fischer instruments.					
Instrument properties	<ul> <li>Ideal for onsite applications due to the col durable instrument design</li> </ul>	<ul> <li>Ideal for onsite applications due to the compact size, the light weight and the robust and durable instrument design</li> </ul>				
	<ul> <li>Intuitive operation of the menu navigation matically, like a smart phone</li> </ul>	• Intuitive operation of the menu navigation and graphic display. The display turns auto				
	<ul> <li>Second display for reading the measuren instrument, e.g., for measuring overhead</li> <li>Different languages are selectable</li> </ul>	nent results directly on the top side of the				
	Manufacturer's certificate, included in the	e scope of supply				
Generating measurements	<ul> <li>The specimen's shape and permeability has measurement results</li> </ul>	nave a comparatively low influence on the				
	<ul> <li>Patented conductivity compensation for m materials</li> </ul>	<ul> <li>Patented conductivity compensation for measurements on non-magnetic substrate materials</li> </ul>				
	<ul> <li>Two special measuring modes in accorded IMO PSPC (90/10-Rule) and SSPC-PA2</li> </ul>	ance with the measurement regulations				
Applications	Steel or iron substrates (Fe)	Nonferrous metal substrates (NF)				
Examples	<ul> <li>Zinc, chromium, copper, paint, varnish and plastic coatings on steel, iron or</li> </ul>	<ul> <li>Paint, varnish or plastic coatings on aluminium, copper or brass</li> </ul>				
	cast iron (Fe)	<ul> <li>Anodized coatings on aluminium</li> </ul>				
	The instrument is particularly suited for measurements on thick metal coatings (e. g. 300 µm/11.8 mils copper) and thick protective coatings (e. g. 5 mm/197 mils enamel) on steel and iron.					
Evaluation						
Statistics	Display of mean value, standard deviation, block	MIN, MAX and number of measurements per				
PC software included in the scope of supply	PC software FISCHER DataCenter with the following functionality: Transferring and archiving measurement data, comprehensive statistical and graphical evaluations, easy creation and printing of inspection reports					
Measuring Modes						
Standard (Std)	Standard measuring mode for simple, universal coating thickness measurements, all mea urement functions are available.					
IMO PSPC 90/10 (90.10)	90/10 rule stored in the instrument for coating thickness measurements according to the requirements of the "Performance Standard for Protective Coatings" of the International Maritime Organization (IMO PSPC).					

for Protective Coatings (SSPC).

Coating thickness measurement according to the test specification SSPC-PA2 of the Society

SSPC-PA2 (SSPC)

Block size Adjustable between 2 and 20 single readings per block

Tolerance limits Adjustable, depending on the selected measuring mode

Offset value In the standard mode, the freely adjustable offset value is deducted automatically from the

measured value. Thus, one obtains the thickness of the top coating if for instance the

interim coating is known.

Units of measurement Selectable µm or mils

Continuous display mode Measurement in "continuous display mode" for continuous sampling of the surfaces, e.g.,

in the manufacture of tanks and containers.

Normalization Adaptation to the substrate material and the shape of the specimen.

Factory calibration Each individual instrument is factory calibrated at several reference points with the

greatest care to ensure the highest possible degree of trueness.

Corrective calibration (Adjustment)

Adaptation to the substrate material and the shape of the specimen and to a thickness

value using a calibration foil.

Simple Calibration

Adaption to the coating and substrate material in one step using a coated reference part with a coating thickness higher than 200 µm (7.87 inches). Nevertheless, this kind of calibration supplies only a lower accuracy as specified in the sections Trueness and

Repeatability Precision.

#### **General Features**

Calibration

Measuring method Magnetic method (ISO 2178, ASTM D7091, Measurement of non-magnetic coatings on

magnetic substrates);

Eddy current method (ISO 2360, ASTM D7091, Measurement of non-conductive coatings

on non-magnetic substrate metals);

Automatic selection of the measuring method corresponding to the substrate material

Probe Probe tip radius: 1.2 mm (46.8 mils); Probe tip material: Ruby

Max. 10,000 individual readings; the contents of the memory is retained even without Data memory

batteries

Measuring frequency More than 70 measurements per minute

Measurement acquisition Automatic upon placement of the probe; indication of the measurement with a beep

visually with a green lit LED

Display limit value violation Acoustically through 2 short beeps and visually with a red lit LED

Display • Graphic display with an automatically turning display in order to read the measurement

results in many different instrument positions

• LCD display on the top side of the instrument, e.g., for reading the measurement value

for measuring overhead

Languages Many different display languages are selectable: German, English and several other

European and Asian languages

USB port 2.0 compatible, mini type B socket, for connecting a PC

Data transfer Single readings, mean values, group separator

2 Batteries, LR6, AA, 1.5 V Power supply

DUALSCOPE® MPORH-FP 17

## DUALSCOPE® MPORH-FP

Admissible ambient temperature

range during operation Weight (incl. batteries)

Dimensions (W  $\times$  D  $\times$  H)

0 ... +40 °C (+32 ... +104 °F)


184 g (6.5 oz)

Width: 64 mm (2.5 "); Depth: 28 mm (1.1 "); Height: 85 mm (3.35 ")

#### **Dimensions**

Instrument Width: 64 mm (2.52 "); Depth: 28 mm (1.10 "); Height: 85 mm (3.35 ")

Probe of instrument MPOR-FP



A 4	easu					
$\mathbf{n}$	DOCLI	rom	Ont	~	an	CO
	IEU SU	ICIII	CIII		ш	uc

#### Steel or iron substrates (Fe)

#### Nonferrous metal substrates (NF)

0 ... 7000 µm (273 mils)

0 ... 2500 µm (97.5 mils)

#### **Trueness**

#### based on factory calibration standards of the Helmut Fischer GmbH

Steel or iron substrates (Fe)

 $0 \dots 150 \ \mu m \le 5 \ \mu m$  $150 \dots 3000 \, \mu m \le 3 \% \, of \, reading$  $3000 \dots 6000 \, \mu m$ :  $\leq 5 \%$  of reading

... 5.85 mils:  $\leq 0.195 \text{ mils}$  $5.85 \dots 117 \text{ mils}$ :  $\leq 2 \% \text{ of reading}$ 117 ... 234 mils:  $\leq 5$  % of reading Nonferrous metal substrates (NF) 0 ...  $50 \, \mu m$ :  $\leq 1 \, \mu m$ 

50 ... 1000  $\mu m$ :  $\leq 2 \%$  of reading  $1000 \dots 2200 \ \mu m \le 3 \% \ of \ reading$ 

 $0 \dots 2 \text{ mils} \le 0.039 \text{ mils}$  $2 \dots 39 \text{ mils}$ :  $\leq 2 \% \text{ of reading}$ 

39 ... 85.8 mils:  $\leq$  3 % of reading

#### **Repeatability Precision**

based on factory calibration standards of the Helmut Fischer GmbH, 5 single measurement readings on each standard

### Steel or iron substrates (Fe)

 $0 \dots 200 \, \mu m \le 2 \, \mu m$  $200 \dots 6000 \, \mu m \le 1 \% \, of \, reading$ 

 $0 \dots 7.8 \text{ mils} \le 0.078 \text{ mils}$ 7.8 ... 234 mils:  $\leq 1 \%$  of reading

#### Nonferrous metal substrates (NF)

0 ...

 $50 \ \mu m$ :  $\leq 0.5 \ \mu m$ 50 ... 1000  $\mu m$ :  $\leq 1 \%$  of reading 1000 ... 2200  $\mu m$ :  $\leq 1.5$  % of reading

2 mils:  $\leq 0.0195 \text{ mils}$ 0 ... 2 ... 39 mils:  $\leq$  1 % of reading  $39 \dots 85.8 \text{ mils} \le 1.5 \% \text{ of reading}$ 

#### **Ordering Data**

605-115

DUALSCOPE MPORH-FP, probe with cable (80 cm; 31.5 ") permanently connected to the instrument

#### Scope of Supply

Instrument case; protective instrument cover; lanyard; 2 batteries; metal plates NF/FE and ISO/NF for testing purposes; calibration foil (foil thickness about 75 µm und 500 µm (2.95 and 19.7 inches)); operator's manual; manufacturer's certificate; USB cable; support CD with USB drivers, software program FISCHER DataCenter for convenient evaluating, documenting and archiving of the measurement data, software program PC-Datex for exporting the measurement data to an Excel spreadsheet

DUALSCOPE® is a registered trademark of Helmut Fischer GmbH Institut für Elektronik und Messtechnik in Germany and in other coun-









Helmut Fischer AG und Helmut Fischer Technologie AG CH-6331 Hünenberg, Switzerland



IfG-Institute for Scientific Instruments GmbH 12489 Berlin, Germany

Fischer Instrumentation (GB) Ltd Lymington, Hampshire SO41 8JD, England



Fischer Technology, Inc. Windsor, CT 06095, USA



Helmut Fischer S. de R.L. de C.V. 76230 Querétaro, QRO, Mexico Fischer Instrumentation Electronique 78180 Montigny le Bretonneux, France

Helmut Fischer S.R.L. 20099 Sesto San Giovanni (Milano), Italy

Fischer Instruments, S.A. 08018 Barcelona, Spain

Helmut Fischer Meettechniek B.V. 5627 GB Eindhoven, The Netherlands

Fischer do Brasil 04561-001 São Paulo, Brazil

Fischer Instrumentation (Taiwan) Ltd Taipei City 11493, Taiwan

Fischer Instruments K.K. Saitama-ken 340-0012, Japan

Nantong Fischer Instrumentation Ltd Shanghai 200333, P.R. China



Fischer Instrumentation (Far East) Ltd Kwai Chung, N.T., Hong Kong

Fischer Measurement Technologies (India) Pvt. Ltd Pune 411036, India

Fischer Instrumentation (S) Pte Ltd Singapore 658065, Singapore

Helmut Fischer Korea Co., Ltd Seoul-City, Republic of Korea

Fischer Technology (M) SDN Bhd 47301 Petaling Jaya, Malaysia

Helmut Fischer Thailand Co., Ltd Bangkok 10250, Thailand

Fischer Instruments Middle East FZE P.O.Box Dubai 371100, United Arab Emirates







